



Certificate of Calibration

DATE CALIBRATED: 11/26/2021

SITE NAME: Allied Concrete

METER TYPE AND MODEL: pH / pHTestr 30

METER SERIAL NUMBER: 191 - Double Junction

PROBE SERIAL NUMBER: N/A

THERMOMETER NUMBER: 1728405

NIST TEMP READING:	<u>4.0°C</u>	<u>12.8°C</u>	<u>29.5°C</u>
METER TEMP READING:	<u>3.9°C</u>	<u>12.8°C</u>	<u>28.4°C</u>
CORRECTION FACTOR:	<u>+0.1°C</u>	<u>0°C</u>	<u>+1.1°C</u>

SIGNATURE OF CALIBRATOR: *Pem m...*

SO # 169629